



Q606 Calibration System Requirements (Rev 12/31/14)

The Seller shall provide and maintain a calibration system that complies with one or more of the following Calibration System Requirements standards: ISO/IEC 17025, ISO 10012:2003, or NCSL Z540.3. All calibrated equipment shall be traceable to the National Institute of Standards and Technology (NIST).

Compliance with the provisions of this clause in no way relieves the Seller of the final responsibility to furnish acceptable supplies or services as specified herein.

This provision shall be applicable to all inspection, test and measuring equipment supplied by the buyer for the use of the Seller, as well as the Seller's own equipment.

Seller shall flow the substance of this clause to its subcontractors, including this sentence, changing the relationship of the parties as appropriate to preserve the intent of the words.