

# GENERAL DYNAMICS

Ordnance and Tactical Systems

## Seattle Supplier Quality Requirements

Quality Note
[4.34]

Revision	Date
-	05/16/2012

## CERTIFIED CALIBRATION REPORT

### REVISION HISTORY

REV.	EFFECTIVE DATE	NOTES
-	05/16/2012	Initial Release

The applicable revision of this document is determined by:

- The revision specified on the purchase order, or
- The revision in effect at the time of the purchase order if no revision is listed on the purchase order.

### REQUIREMENTS

A certified calibration report, which must include the date and signature of a representative of the supplier is required. The supplier must identify standards used that are traceable to the National Institute of Standards and Technology (NIST).