

1. SQAR Number	SQAR-12
2. SQAR Revision	8
3. SQAR Title	Calibration System and Report
4. SQAR Detail Description	
<p>The Supplier shall have a robust calibration process containing all inspection equipment utilized to determine the acceptance/rejection of parts. The Supplier shall provide upon request a certified calibration report for each instrument. The report shall identify NIST traceable standards and show actual values measured. This is flowed for those suppliers that meet SQAR-3; as this requirement is already in effect for suppliers that meet SQAR-1 and SQAR-2.</p>	
5. Disclaimer	
<p>The following Supplemental Quality Assurance Requirement (SQAR) clause is a requirement when specified by number within the Request for Quote (RFQ), Purchase Order (PO), or Quality Assurance Document (QAD). In the event of conflict between the General Supplier Quality Assurance Requirements document, PAL 164, and a SQAR flowdown, the SQAR will always supersede. Flowdown of this SQAR to sub-tier suppliers is required to ensure final product compliance.</p>	